


manufacturers

Declaration Of Conformity



Product identification

Product : The Non-Automatic Weighing Instrument.
Brand : CAS Corporation.
Type : **AP-1** Version : 00.

Means of conformity

The product is in conformity with following Directives based on test results using the following harmonized standards.

Testing

- EMC (Directive : 89/336/EEC)

Carried out by : NMI P. O. BOX 15 9822 ZG Niekerk
Standards used : EN 50081-1:1991
NMI Project No. : 10018373

- LVD (Directive : 93/68/EEC)

Carried out by : SKY ENGINEERING CO., LTD.
5FL Soosung BLDG, 462-1, Amsa-Dong, Kangdong-Ku, Seoul, Korea.
Standards used : EN 60950:1992 + A1:1993 + A2:1993 + A3 : 1995 +A4:96
Test Report No. : SE-ETS-981229-02.

Manufacturer : CAS Corporation

#19, Ganap-Ri, Kwangjuk-Mean, Kyeonggi-Do, Rep of KOREA.

Signature

Name : Sang-Zee Lee.
Position : Director.
Date : 21 April, 1999.

A handwritten signature in black ink, followed by the date '4/21/99' written in a similar style.



Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N° R76/1992-NL-02.09

Project number 200248

Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: P.P.M. van Enckevort

Applicant

Name: CAS Corporation
Address: #19 Kanap-Ri, Kwangjuk-Myun, Yangju-Gun, Kyungki-Do
Korea

Manufacturer of the certified pattern

Name: CAS Corporation
Address: #19 Kanap-Ri, Kwangjuk-Myun, Yangju-Gun, Kyungki-Do
Korea

Identification of the certified pattern

Type : AP series

$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$

$e \geq 1 \text{ g}$

$n \leq 3000$ divisions for instruments with a multi-interval, per partial weighing range, with a maximum of two weighing ranges.

$T \leq - \text{Max}$

Temperature range: $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the type-approval certificate and the description with number T6027 and the appertaining documentation folder), with the requirements of the following Recommendation of the International Organisation of Legal Metrology (OIML):

R76

edition 1992

for accuracy class **III**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht
Telephone +31 78 6332332
Telefax +31 78 6332309

NMI B.V.
(Chamber of Commerce no.27.228.701)

Subsidiary companies:
NMI Van Swinden Laboratorium B.V. (27228703)
NMI Certin B.V. (27.233.418)
Verispect B.V. (27.228.700)

This document is issued under the provision that NMI B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission.



Nederlands Meetinstituut

O Member State
The Netherlands

OIML Certificate N° R76/1992-NL-02.09
Project number 200248
Page 2 of 2

The conformity was established by tests described in the associated test report:
N° R76/1992-NL-02.09, that includes 54 pages.

The issuing authority
P.P.M. van Enckevort
Manager Certification Delft

b/a

16 April 2002

The OIML member
G.J. Faber

16 April 2002

*
**

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.



Nederlands Meetinstituut

EC type-approval certificate

Number **T6027** revision 0

Project number 200248

Page 1 of 4

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Notified Body Number 0122

In accordance with The Council Directive 90/384/EEC on non-automatic weighing instruments.

Applicant CAS Corporation
#19 Kanap-Ri, Kwangjuk-Myun, Yangju-Gun, Kyungki-Do
Korea

In respect of A class **III**, electronic, multi-interval **non-automatic weighing instrument**,
intended to be used for direct sales to the public.
Manufacturer : CAS Corporation
Type : AP series

Characteristics $n \leq 3000$ divisions (per partial weighing range)
 $6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
 $e \geq 1 \text{ g}$
maximum of two partial weighing ranges

Temperature range $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$

In the description number T6027 revision 0 further characteristics are described.

Valid until 16 April 2012

Description and documentation The instrument is described in the description number T6027 revision 0 and documented in the documentation folder T6027-1, appertaining to this EC type-approval certificate.

Delft, 16 April 2002
NMI Certin B.V.

b/a

P.P.M. van Enckevort
Manager Certification Delft


Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht
Telephone +31 78 6332332
Telefax +31 78 6332309

NMI B.V.
(Chamber of Commerce no.27.228.701)

Subsidiary companies:
NMI Van Swinden Laboratorium B.V. (27228703)
NMI Certin B.V. (27.233.418)
Verispect B.V. (27.228.700)

This document is issued under the provision that NMI. B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission.



1 General information about the non-automatic weighing instrument

All properties of the non-automatic weighing instrument, whether mentioned or not, may not be in conflict with the legislation.

1.1 Essential parts

The electronics

The mechanical assembly with load cell.

EMC protection measures:

- The A/D board is shielded with a metal cover.

1.2 Essential characteristics

Power supply: 110 – 120 or 220 - 240 V AC 50/60 Hz.

1.3 Essential shapes

The non-automatic weighing instrument is built according to drawing "THE EXPLODED VIEW", drawing number 1.

The data plate is secured against removal by sealing or will be destroyed when removed.

To secure components that may not be dismantled or adjusted by the user, the non-automatic weighing instrument has to be secured in a suitable manner on the locations indicated in the drawings "SEALING METHOD", drawing numbers 2 and 3.

The securing component has to bear either:

- a mark of the manufacturer laid down in a notified body approved quality system (Annex II of the directive 90/384/EEC), or
- an official mark of a Member State of the EEC, or an other party to the EEA agreement.

Inside the cabinet is a calibration lock, located on the main board.

1.4 Conditional parts

The non-automatic weighing instrument may be equipped with peripheral equipment which is used for the applications listed in article 1(2)(a) of the EC Directive (90/384/EEC), if the peripheral equipment is certified to be connected to an EC type-approved non-automatic weighing instrument by a Notified Body appointed to certify non-automatic weighing instruments according to paragraph I of Annex II of the EC directive on Non-Automatic Weighing Instruments.

The non-automatic weighing instrument may be equipped with an Electronic Point of Sale (EPoS) or an Electronic Cash Register (ECR), if these EPoS and ECR are certified to be connected to an EC type-approved non-automatic weighing instrument by a Notified Body appointed to certify non-automatic weighing instruments according to paragraph I of Annex II of the EC Directive on Non-Automatic Weighing Instruments.

The non-automatic weighing instrument may be equipped with Electronic Funds Transfer equipment (EFT/ECU), if these EFT/ECU represent only the price total on the display.

If the non-automatic weighing instrument is liable to be tilted the instrument must be equipped with a level indicator with a sensitivity of at least 2 mm for a tilt of 2/1000.

2 Information about the main constituent parts of the non-automatic weighing instrument

2.1 The electronics

2.1.1 Essential parts

Description	Drawing number	Rev.	Remarks
PARTS LOCATION	4	--	--
ANALOG MODULE	6144-M00-0100	--	--
DIGITAL MODULE	5	--	--
AP SERIES FULL PART LIST	--	--	2-pages

2.1.2 Essential characteristics

List of devices:

- determination stability of equilibrium;
- zero indicator;
- semi-automatic zero-setting;
- initial zero-setting;
- zero-tracking;
- semi-automatic subtractive tare balancing;
- tare weighing;
- preset tare;
- indication of stable equilibrium;
- calibration / set-up mode via a switch on the main board;
- acting upon significant faults;
- checking the display;
- price calculation;

2.1.3 Essential shapes

See the drawing: "THE EXPLODED VIEW", drawingnumber 1.


2.1.4 Conditional parts

The interface section is located on a separate interface board. The non-automatic weighing instrument may be equipped with the following protective interface that has not to be secured:

- RS232;

2.1.5 Non-essential parts

Display;
 Keyboard;



2.2 The mechanical assembly with load cell

2.2.1 Essential parts

Description	Drawing number	Rev.	Remarks
THE EXPLODED VIEW	1	--	--

2.2.2 Essential characteristics

$e_1 \geq E_{\max}/10000$;
Excitation power supply 12.25 V DC.

3 Approval conditions

See chapter 1.3, essential shapes.

4 Seals and verification marks

See chapter 1.3, essential shapes.

5 CE-mark of conformity and inscriptions

The marks, facilities for the marks and the inscriptions on the non-automatic weighing instrument fulfill the requirements of article 1 of Annex IV.