

Manufacturer's Declaration of Conformity

Product Identification

Product Description : Non-Automatic Weighing Instrument

Manufactured by : CAS Corporation

Type : ER^{JR}

Year in which the CE-mark was affixed : 2007


Version : 01

EC-TAC Number : UK2829

Serial Number :

Means of conformity

The product is in conformity with following Directives based on test results using the following harmonized standards.



Testing for CE marking

- EMC (Directive : 89/336/EEC)

Carried out by : HCT CO., LTD

San 136-1, Ami-Ri, Bubal-Eup, Icheon-Si, Kyoungki-Do, Korea

Standards used : EN 61326:1997+A1:1998+A2:2001+A3:2003,

EN 61000-3-2:2000,

EN 61000-3-3:1995+A1:2001

EN 61000-4-2:1995+A1:1998+A2:2001 EN 61000-4-3:2002+A1:2002

EN 61000-4-4:1995+A1:2001,

EN 61000-4-5:1995+A1:2001

EN 61000-4-6:1996+A1:2001,

EN 61000-4-11:2004

Test Report No. : HCT-C07-0806

- LVD (Directive : 73/23/EEC)

Adaptor Model Name : HAPU05XY ("X" can be A, B, C, E, F, H, J, M, X, Y, or Z; "Y" can be 1, 2, 3, 4, 5, or 6.)

Carried out by : TUV Product Service GmbH

Ridlerstr, 65, D-80339, Munchen, Germany

Standards used : IEC 60950-1:2001, First Edition

Test Report Ref No. : 081-50824-000

- NAWI (Directive : 90/384/EEC)

Carried out by : National Weights & Measures Laboratory

Stanton Avenue, Teddington, Middlesex, TW11 0JZ, United Kingdom

Standards used : EN 45501

Test Report No. : SN 1025, SN 1027, SN 1029

Maxin Road 448, Xinqiao Town, Songjian District, Shanghai, China

Name : He Chun Rong

Signature :

A handwritten signature in black ink, appearing to read 'He Chun Rong', written over a faint circular stamp.

Position : Director of Quality Assurance

Date : JAN . 28 . 2008